

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087231.D
 Acq On : 27 Jun 2025 18:25
 Operator : JC\MD
 Sample : Q2414-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jun 27 23:49:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	186806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	348147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	143304	53.306	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.620%
35) Dibromofluoromethane	8.177	113	106715	47.412	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.820%
50) Toluene-d8	10.570	98	436488	46.888	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.780%
62) 4-Bromofluorobenzene	12.847	95	144331	44.682	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.360%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	44924	22.383	ug/l	99
18) Methyl Acetate	5.041	43	25808	7.100	ug/l	96
20) Methylene Chloride	5.294	84	10324	2.569	ug/l	94
37) Ethyl Acetate	7.577	43	15289	1.367	ug/l	98
43) Isopropyl Acetate	8.694	43	10771	1.777	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087231.D
Acq On : 27 Jun 2025 18:25
Operator : JC\MD
Sample : Q2414-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jun 27 23:49:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration

