

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087214.D
 Acq On : 27 Jun 2025 12:29
 Operator : JC\MD
 Sample : Q2418-01 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VAC-TRUCK-4074

Quant Time: Jun 27 23:42:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	160109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	252544	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128959	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	120855	52.452	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.900%
35) Dibromofluoromethane	8.171	113	96593	59.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.320%
50) Toluene-d8	10.565	98	342358	50.699	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.400%
62) 4-Bromofluorobenzene	12.847	95	130536	55.709	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.420%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087214.D
Acq On : 27 Jun 2025 12:29
Operator : JC\MD
Sample : Q2418-01 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VAC-TRUCK-4074

Quant Time: Jun 27 23:42:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. -0.000 min
Lab File: VN087214.D
Acq: 27 Jun 2025 12:29




Instrument :
MSVOA_N
ClientSampleId :
VAC-TRUCK-4074


Tgt Ion:168 Resp: 160109
Ion Ratio Lower Upper
168 100
99 60.2 47.9 71.9



#33
1,2-Dichloroethane-d4
Concen: 52.452 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. -0.000 min
Lab File: VN087214.D
Acq: 27 Jun 2025 12:29

Tgt Ion: 65 Resp: 120855
Ion Ratio Lower Upper
65 100
67 54.9 0.0 108.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN087214.D

Acq: 27 Jun 2025 12:29

Instrument :

MSVOA_N

ClientSampleId :

VAC-TRUCK-4074





Tgt Ion:114 Resp: 252544

Ion Ratio Lower Upper

114 100

63 19.7 0.0 42.0

88 16.2 0.0 32.0



#35

Dibromofluoromethane

Concen: 59.161 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN087214.D

Acq: 27 Jun 2025 12:29




Tgt Ion:113 Resp: 96593


Ion Ratio Lower Upper

113 100

111 104.0 82.6 124.0

192 17.4 13.5 20.3









#50
Toluene-d8
Concen: 50.699 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN087214.D
Acq: 27 Jun 2025 12:29




Instrument :
MSVOA_N
ClientSampleId :
VAC-TRUCK-4074


Tgt Ion: 98 Resp: 342358
Ion Ratio Lower Upper
98 100
100 65.3 52.7 79.1



#62
4-Bromofluorobenzene
Concen: 55.709 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN087214.D
Acq: 27 Jun 2025 12:29

Tgt Ion: 95 Resp: 130536
Ion Ratio Lower Upper
95 100
174 70.6 0.0 142.2
176 66.6 0.0 134.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN087214.D
Acq: 27 Jun 2025 12:29

Instrument :
MSVOA_N
ClientSampleId :
VAC-TRUCK-4074

Tgt Ion:117 Resp: 283675
Ion Ratio Lower Upper
117 100
82 57.1 45.0 67.4
119 31.1 25.9 38.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN087214.D
Acq: 27 Jun 2025 12:29

Tgt Ion:152 Resp: 128959
Ion Ratio Lower Upper
152 100
115 60.9 30.6 91.8
150 158.4 0.0 347.8

