

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049583.D
 Acq On : 28 Jun 2018 12:14
 Operator : MD\SY
 Sample : VN0628WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0628WBL01

Quant Time: Jun 29 02:26:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1448406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2312457	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1904751	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	732532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	941822	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
35) Dibromofluoromethane	7.59	113	839461	44.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.74%	
50) Toluene-d8	10.09	98	3119219	43.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.58%	
62) 4-Bromofluorobenzene	12.40	95	886921	37.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.30%	
Target Compounds						
16) Acetone	3.82	43	98119	21.04	ug/l	Qvalue 92
18) Methyl Acetate	4.33	43	37900	2.66	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049583.D
Acq On : 28 Jun 2018 12:14
Operator : MD\SY
Sample : VN0628WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0628WBL01

Quant Time: Jun 29 02:26:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

