

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049586.D
 Acq On : 28 Jun 2018 13:41
 Operator : MD\SY
 Sample : PB110627ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110627ZHE#09

Quant Time: Jun 29 02:33:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1170416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1892973	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1591538	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	640767	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	791334	50.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.96%	
35) Dibromofluoromethane	7.59	113	636012	41.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.06%	
50) Toluene-d8	10.09	98	2546718	43.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.34%	
62) 4-Bromofluorobenzene	12.40	95	771109	39.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.98%	

Target Compounds						Qvalue
16) Acetone	3.82	43	83321	22.11	ug/l	93
18) Methyl Acetate	4.33	43	155919	13.57	ug/l	94
20) Methylene Chloride	4.55	84	33982	2.16	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049586.D
Acq On : 28 Jun 2018 13:41
Operator : MD\SY
Sample : PB110627ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110627ZHE#09

Quant Time: Jun 29 02:33:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

