

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049587.D
 Acq On : 28 Jun 2018 14:07
 Operator : MD\SY
 Sample : PB110627ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110627ZHE#10

Quant Time: Jun 29 03:08:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1159714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1878985	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1551136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	601098	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	769731	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	7.59	113	641634	42.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.42%	
50) Toluene-d8	10.09	98	2510583	43.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.74%	
62) 4-Bromofluorobenzene	12.40	95	718993	37.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.14%	
Target Compounds						
16) Acetone	3.82	43	74892	20.05	ug/l	95
18) Methyl Acetate	4.33	43	158145	13.89	ug/l	96
20) Methylene Chloride	4.55	84	35808	2.30	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049587.D
Acq On : 28 Jun 2018 14:07
Operator : MD\SY
Sample : PB110627ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110627ZHE#10

Quant Time: Jun 29 03:08:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

