

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049597.D
 Acq On : 28 Jun 2018 18:25
 Operator : MD\SY
 Sample : J3715-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS-CURB-CON

Quant Time: Jun 29 03:36:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1105775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1791758	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1466983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	591650	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	743311	50.68	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.36%	
35) Dibromofluoromethane	7.59	113	664529	45.84	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.68%	
50) Toluene-d8	10.09	98	2399622	43.47	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	86.94%	
62) 4-Bromofluorobenzene	12.40	95	682551	37.40	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	74.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	7959	1.05	ug/l	99
16) Acetone	3.82	43	76714	21.54	ug/l	96
18) Methyl Acetate	4.33	43	165859	15.28	ug/l	97
20) Methylene Chloride	4.55	84	3312473	222.72	ug/l	90
25) 2-Butanone	6.85	43	11167	2.34	ug/l #	90
84) 1,2,4-Trimethylbenzene	13.04	105	69236	1.71	ug/l	99
95) Naphthalene	15.14	128	140417	10.44	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049597.D
Acq On : 28 Jun 2018 18:25
Operator : MD\SY
Sample : J3715-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS-CURB-CON

Quant Time: Jun 29 03:36:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

