

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049600.D
 Acq On : 28 Jun 2018 19:43
 Operator : MD\SY
 Sample : J3048-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-S

Manual Integrations
 APPROVED

MMDadoda
 6/29/2018 1:46:44 PM

Quant Time: Jun 29 03:47:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1069906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1750854	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1438223	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	583631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	736458	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
35) Dibromofluoromethane	7.59	113	644434	45.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.98%	
50) Toluene-d8	10.09	98	2325688	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
62) 4-Bromofluorobenzene	12.40	95	680779	38.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.34%	
Target Compounds						
8) Diethyl Ether	3.40	74	9839m	1.34	ug/l	Qvalue
16) Acetone	3.82	43	109960	31.91	ug/l	98
18) Methyl Acetate	4.33	43	138812	13.21	ug/l	93
20) Methylene Chloride	4.55	84	1307605	90.87	ug/l	91
25) 2-Butanone	6.84	43	22724	4.91	ug/l	94
95) Naphthalene	15.14	128	300475	19.73	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049600.D
Acq On : 28 Jun 2018 19:43
Operator : MD\SY
Sample : J3048-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-S

Manual Integrations
APPROVED

MMDadoda
6/29/2018 1:46:44 PM

Quant Time: Jun 29 03:47:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

