

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049605.D
 Acq On : 28 Jun 2018 21:52
 Operator : MD\SY
 Sample : J3242-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-062618-C

Quant Time: Jun 29 04:13:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1152427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1857153	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1535698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	633707	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	769793	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
35) Dibromofluoromethane	7.59	113	678475	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
50) Toluene-d8	10.09	98	2468014	43.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.28%	
62) 4-Bromofluorobenzene	12.40	95	741990	39.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.44%	
Target Compounds						
16) Acetone	3.82	43	82927	22.34	ug/l	Qvalue # 86
18) Methyl Acetate	4.34	43	159608	14.10	ug/l	95
20) Methylene Chloride	4.55	84	216999	14.00	ug/l	90
25) 2-Butanone	6.84	43	10871	2.18	ug/l	95
68) m/p-Xylenes	11.62	106	56796	2.31	ug/l	94
84) 1,2,4-Trimethylbenzene	13.04	105	59956	1.38	ug/l	97
95) Naphthalene	15.14	128	187497	12.44	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049605.D
Acq On : 28 Jun 2018 21:52
Operator : MD\SY
Sample : J3242-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-062618-C

Quant Time: Jun 29 04:13:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

