

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049613.D
 Acq On : 29 Jun 2018 2:09
 Operator : MD\SY
 Sample : J3242-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-062618-F

Manual Integrations
 APPROVED

MMDadoda
 6/29/2018 1:46:51 PM

Quant Time: Jun 29 04:54:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1103688	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1760415	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1458946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	587841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	719070	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
35) Dibromofluoromethane	7.59	113	647400	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
50) Toluene-d8	10.09	98	2356667	43.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.92%	
62) 4-Bromofluorobenzene	12.40	95	685901	38.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.50%	
Target Compounds						
16) Acetone	3.82	43	64032	18.02	ug/l	# 87
18) Methyl Acetate	4.33	43	144566	13.34	ug/l	94
20) Methylene Chloride	4.55	84	174383	11.75	ug/l	93
25) 2-Butanone	6.85	43	10335m	2.17	ug/l	
95) Naphthalene	15.14	128	103370	8.31	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049613.D
Acq On : 29 Jun 2018 2:09
Operator : MD\SY
Sample : J3242-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-062618-F

Manual Integrations
APPROVED

MMDadoda
6/29/2018 1:46:51 PM

Quant Time: Jun 29 04:54:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

