

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049616.D
 Acq On : 29 Jun 2018 3:26
 Operator : MD\SY
 Sample : PB110667TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110667TB

Quant Time: Jun 29 05:04:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1150554	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1864484	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1536157	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	561602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	759361	49.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.52%	
35) Dibromofluoromethane	7.59	113	637276	42.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.50%	
50) Toluene-d8	10.09	98	2492283	43.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.78%	
62) 4-Bromofluorobenzene	12.40	95	716176	37.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.42%	
Target Compounds						
16) Acetone	3.82	43	61485	16.59	ug/l	93
18) Methyl Acetate	4.33	43	131680	11.66	ug/l	95
20) Methylene Chloride	4.55	84	19678	1.27	ug/l #	79
95) Naphthalene	15.14	128	27780	3.86	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049616.D
 Acq On : 29 Jun 2018 3:26
 Operator : MD\SY
 Sample : PB110667TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110667TB

Quant Time: Jun 29 05:04:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

