

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049618.D
 Acq On : 29 Jun 2018 4:17
 Operator : MD\SY
 Sample : PB110663TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110663TB

Quant Time: Jun 29 08:46:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1042488	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1702895	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1400471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	521533	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	682910	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
35) Dibromofluoromethane	7.59	113	567433	41.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.38%	
50) Toluene-d8	10.09	98	2263670	43.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.30%	
62) 4-Bromofluorobenzene	12.40	95	643784	37.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.24%	
Target Compounds						
16) Acetone	3.83	43	40153	11.96	ug/l	94
18) Methyl Acetate	4.33	43	72440	7.08	ug/l #	90
20) Methylene Chloride	4.54	84	16721	1.19	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049618.D
Acq On : 29 Jun 2018 4:17
Operator : MD\SY
Sample : PB110663TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110663TB

Quant Time: Jun 29 08:46:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

