

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062818\
 Data File : VN049619.D
 Acq On : 29 Jun 2018 4:43
 Operator : MD\SY
 Sample : J3612-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-4(3-3.5)

Quant Time: Jun 29 08:52:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1133678	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1811119	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1493079	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	564981	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	739212	49.16	ug/l	0.00
Spiked Amount			Recovery	=	98.32%	
35) Dibromofluoromethane	7.59	113	500694	34.17	ug/l	0.00
Spiked Amount			Recovery	=	68.34%	
50) Toluene-d8	10.09	98	2425104	43.47	ug/l	0.00
Spiked Amount			Recovery	=	86.94%	
62) 4-Bromofluorobenzene	12.40	95	690794	37.45	ug/l	0.00
Spiked Amount			Recovery	=	74.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	58828	16.11	ug/l	96
18) Methyl Acetate	4.34	43	74614	6.70	ug/l	99
20) Methylene Chloride	4.55	84	22131	1.45	ug/l	95
64) Tetrachloroethene	10.63	164	44394	2.90	ug/l	96
95) Naphthalene	15.14	128	9262	2.67	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062818\
Data File : VN049619.D
Acq On : 29 Jun 2018 4:43
Operator : MD\SY
Sample : J3612-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-4(3-3.5)

Quant Time: Jun 29 08:52:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

