

Data File : VN049622.D
Acq On : 29 Jun 2018 6:00
Operator : MD\SY
Sample : J3612-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-2(8.5-9)

Manual Integrations
APPROVED

MMDadoda
6/29/2018 1:46:50 PM

Quant Time: Jul 09 03:38:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1180067	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1907465	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1582220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	581579	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	772779	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
35) Dibromofluoromethane	7.59	113	102358	6.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	13.26%	
50) Toluene-d8	10.09	98	2570767	43.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.50%	
62) 4-Bromofluorobenzene	12.40	95	736872	37.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.86%	
Target Compounds						
16) Acetone	3.82	43	67512	17.765	ug/l	96
18) Methyl Acetate	4.34	43	38296m	3.305	ug/l	
20) Methylene Chloride	4.54	84	19566	1.233	ug/l	88
95) Naphthalene	15.14	128	8375	2.599	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN049622.D
Acq On : 29 Jun 2018 6:00
Operator : MD\SY
Sample : J3612-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-2(8.5-9)

Manual Integrations
APPROVED

MMDadoda
6/29/2018 1:46:50 PM

Quant Time: Jul 09 03:38:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

