

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056549.D
 Acq On : 28 Jun 2019 12:55
 Operator : JC/SP
 Sample : K3519-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 28 23:31:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50433	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	85797	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	51626	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	7408	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	30177	4.88	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	97.60%
5) Dibromofluoromethane	7.59	113	22650	4.79	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	95.80%
6) Toluene-d8	10.09	98	49634	4.64	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	92.80%
7) 4-Bromofluorobenzene	12.40	95	12318	4.50	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	90.00%


Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	27	0.005 ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062819\
Data File : VN056549.D
Acq On : 28 Jun 2019 12:55
Operator : JC/SP
Sample : K3519-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jun 28 23:31:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 5.000 ug/l

RT: 7.65 min Scan# 1243

Delta R.T. 0.00 min

Lab File: VN056549.D

Acq: 28 Jun 2019 12:55


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 50433

Ion Ratio Lower Upper




168 100

99 51.3 41.9 62.9



Abundance Ion 168.00 (167.70 to 168.70): VN056516.D (-1233) (-)

Ion 99.00 (98.70 to 99.70): VN056516.D (-1233) (-)



#2

Vinyl Chloride

Concen: 0.005 ug/l

RT: 2.17 min Scan# 150

Delta R.T. 0.00 min

Lab File: VN056549.D


Acq: 28 Jun 2019 12:55

Tgt Ion: 62 Resp: 27

Ion Ratio Lower Upper



62 100


64 109.8 26.1 39.1#



Abundance Ion 62.00 (61.70 to 62.70): VN056516.D (-128) (-)

Ion 64.00 (63.70 to 64.70): VN056516.D (-128) (-)









#3
 1,2-Dichloroethane-d4
 Concen: 4.885 ug/l
 RT: 8.01 min Scan# 1264
 Delta R.T. 0.00 min
 Lab File: VN056549.D
 Acq: 28 Jun 2019 12:55




Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion: 65 Resp: 30177
 Ion Ratio Lower Upper
 65 100
 67 54.7 0.0 108.8



#4
 1,4-Difluorobenzene
 Concen: 5.000 ug/l
 RT: 8.58 min Scan# 1297
 Delta R.T. 0.00 min
 Lab File: VN056549.D
 Acq: 28 Jun 2019 12:55

Tgt Ion: 114 Resp: 85797
 Ion Ratio Lower Upper
 114 100
 63 18.8 0.0 37.6
 88 15.6 0.0 31.2









#5
 Dibromofluoromethane
 Concen: 4.789 ug/l
 RT: 7.59 min Scan# 1239
 Delta R.T. 0.00 min
 Lab File: VN056549.D
 Acq: 28 Jun 2019 12:55

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion:113 Resp: 22650
 Ion Ratio Lower Upper
 113 100
 111 101.5 81.7 122.5
 192 21.3 16.9 25.3




Abundance Ion 113.00 (112.70 to 113.70): V
 Ion 111.00 (110.70 to 111.70): V
 Ion 192.00 (191.70 to 192.70): V





#6
 Toluene-d8
 Concen: 4.638 ug/l
 RT: 10.09 min Scan# 1440
 Delta R.T. 0.00 min
 Lab File: VN056549.D
 Acq: 28 Jun 2019 12:55

Tgt Ion: 98 Resp: 49634
 Ion Ratio Lower Upper
 98 100
 100 62.2 49.7 74.5



Abundance Ion 98.00 (97.70 to 98.70): VN
 Ion 100.00 (99.70 to 100.70): VN





#7

4-Bromofluorobenzene

Concen: 4.500 ug/l

RT: 12.40 min Scan# 1773

Delta R.T. 0.00 min

Lab File: VN056549.D

Acq: 28 Jun 2019 12:55

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 95 Resp: 12318

Ion Ratio Lower Upper

95 100

174 85.2 0.0 167.8

176 78.9 0.0 158.0



Abundance Ion 95.00 (94.70 to 95.70): VN056516.D (-1763) (-)

Ion 174.00 (173.70 to 174.70): VN056516.D (-1763) (-)

Ion 176.00 (175.70 to 176.70): VN056516.D (-1763) (-)



12.40

4000

2000

0

Time-->



#8

Chlorobenzene-d5

Concen: 5.000 ug/l

RT: 11.41 min Scan# 1628

Delta R.T. 0.00 min

Lab File: VN056549.D

Acq: 28 Jun 2019 12:55


Tgt Ion: 117 Resp: 51626

Ion Ratio Lower Upper

117 100

82 53.7 43.0 64.4

119 30.9 24.7 37.1



Abundance Ion 117.00 (116.70 to 117.70): VN056516.D (-1617) (-)

Ion 82.00 (81.70 to 82.70): VN056516.D (-1617) (-)

Ion 119.00 (118.70 to 119.70): VN056516.D (-1617) (-)

11.41


30000



20000

10000

0

Time-->





#9

1,4-Dichlorobenzene-d4

Concen: 5.000 ug/l

RT: 13.35 min Scan# 1864

Delta R.T. 0.00 min

Lab File: VN056549.D

Acq: 28 Jun 2019 12:55

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion:152 Resp: 7408

Ion Ratio Lower Upper

152 100

115 54.4 27.0 80.8

150 153.1 0.0 306.6

