

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056557.D
 Acq On : 28 Jun 2019 16:02
 Operator : JC/SP
 Sample : K3519-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 28 23:32:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	48073	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	82277	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	50228	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	7632	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	27330	4.64	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	92.80%
5) Dibromofluoromethane	7.59	113	20533	4.53	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	90.60%
6) Toluene-d8	10.09	98	44717	4.36	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	87.20%#
7) 4-Bromofluorobenzene	12.39	95	11636	4.43	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	88.60%


Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	174	0.036 ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062819\
Data File : VN056557.D
Acq On : 28 Jun 2019 16:02
Operator : JC/SP
Sample : K3519-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jun 28 23:32:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 5.000 ug/l

RT: 7.65 min Scan# 1243

Delta R.T. 0.00 min

Lab File: VN056557.D

Acq: 28 Jun 2019 16:02


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 48073

Ion Ratio Lower Upper




168 100

99 51.7 41.9 62.9



Abundance Ion 168.00 (167.70 to 168.70): VN056557.D (-1225) (-)

Ion 99.00 (98.70 to 99.70): VN056557.D (-1225) (-)



#2

Vinyl Chloride

Concen: 0.036 ug/l

RT: 2.17 min Scan# 150

Delta R.T. 0.00 min

Lab File: VN056557.D


Acq: 28 Jun 2019 16:02

Tgt Ion: 62 Resp: 174

Ion Ratio Lower Upper



62 100

64 70.5 26.1 39.1#



Abundance Ion 62.00 (61.70 to 62.70): VN056557.D (-98) (-)

Ion 64.00 (63.70 to 64.70): VN056557.D (-98) (-)





#3

1,2-Dichloroethane-d4

Concen: 4.641 ug/l

RT: 8.01 min Scan# 1264

Delta R.T. 0.00 min

Lab File: VN056557.D

Acq: 28 Jun 2019 16:02


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 65 Resp: 27330




Ion Ratio Lower Upper

65 100

67 54.7 0.0 108.8



Abundance Ion 65.00 (64.70 to 65.70): VN056557.D (-1246) (-)



#4

1,4-Difluorobenzene

Concen: 5.000 ug/l

RT: 8.58 min Scan# 1297

Delta R.T. 0.00 min

Lab File: VN056557.D

Acq: 28 Jun 2019 16:02


Tgt Ion: 114 Resp: 82277

Ion Ratio Lower Upper



114 100

63 19.1 0.0 37.6

88 15.6 0.0 31.2



Abundance Ion 114.00 (113.70 to 114.70): VN056557.D (-1268) (-)









#5
 Dibromofluoromethane
 Concen: 4.527 ug/l
 RT: 7.59 min Scan# 1239
 Delta R.T. 0.00 min
 Lab File: VN056557.D
 Acq: 28 Jun 2019 16:02




Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion:113 Resp: 20533
 Ion Ratio Lower Upper
 113 100
 111 101.5 81.7 122.5
 192 21.1 16.9 25.3



#6
 Toluene-d8
 Concen: 4.357 ug/l
 RT: 10.09 min Scan# 1440
 Delta R.T. 0.00 min
 Lab File: VN056557.D
 Acq: 28 Jun 2019 16:02

Tgt Ion: 98 Resp: 44717
 Ion Ratio Lower Upper
 98 100
 100 62.9 49.7 74.5





#7

4-Bromofluorobenzene

Concen: 4.432 ug/l

RT: 12.39 min Scan# 1772

Delta R.T. -0.01 min

Lab File: VN056557.D

Acq: 28 Jun 2019 16:02

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 95 Resp: 11636

Ion Ratio Lower Upper

95 100

174 72.7 0.0 167.8




176 64.4 0.0 158.0



Abundance Ion 95.00 (94.70 to 95.70): VN056516.D (-1763) (-)

Ion 174.00 (173.70 to 174.70): VN056516.D (-1763) (-)

Ion 176.00 (175.70 to 176.70): VN056516.D (-1763) (-)



#8

Chlorobenzene-d5

Concen: 5.000 ug/l

RT: 11.41 min Scan# 1628

Delta R.T. 0.00 min

Lab File: VN056557.D

Acq: 28 Jun 2019 16:02


Tgt Ion: 117 Resp: 50228

Ion Ratio Lower Upper

117 100

82 53.3 43.0 64.4



119 31.1 24.7 37.1



Abundance Ion 117.00 (116.70 to 117.70): VN056516.D (-1617) (-)

Ion 82.00 (81.70 to 82.70): VN056516.D (-1617) (-)

Ion 119.00 (118.70 to 119.70): VN056516.D (-1617) (-)





#9

1,4-Dichlorobenzene-d4

Concen: 5.000 ug/l

RT: 13.34 min Scan# 1863

Delta R.T. -0.01 min

Lab File: VN056557.D

Acq: 28 Jun 2019 16:02

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:152 Resp: 7632

Ion Ratio Lower Upper

152 100

115 58.2 27.0 80.8

150 158.8 0.0 306.6



Abundance

Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V

