

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056541.D
 Acq On : 28 Jun 2019 9:20
 Operator : JC/SP
 Sample : VN0628WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0628WBL01

Quant Time: Jun 28 23:36:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	55735	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.59	114	79614	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	53559	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	7370	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	33267	4.87	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	97.40%
5) Dibromofluoromethane	7.58	113	25398	5.79	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	115.80%
6) Toluene-d8	10.09	98	55290	5.57	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	111.40%
7) 4-Bromofluorobenzene	12.40	95	12589	4.96	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	99.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062819\
Data File : VN056541.D
Acq On : 28 Jun 2019 9:20
Operator : JC/SP
Sample : VN0628WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0628WBL01

Quant Time: Jun 28 23:36:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration

