

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056543.D
 Acq On : 28 Jun 2019 10:10
 Operator : JC/SP
 Sample : 0.1ppb LOO
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 0.1ppb LOQ

Quant Time: Jun 28 23:30:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	51487	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	87402	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	53630	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	8119	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	32069	5.08	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	101.60%
5) Dibromofluoromethane	7.59	113	24073	5.00	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	100.00%
6) Toluene-d8	10.09	98	53281	4.89	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	97.80%
7) 4-Bromofluorobenzene	12.39	95	13985	5.01	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	100.20%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	535	0.104	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062819\
Data File : VN056543.D
Acq On : 28 Jun 2019 10:10
Operator : JC/SP
Sample : 0.1ppb LOQ
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
0.1ppb LOQ

Quant Time: Jun 28 23:30:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration

