

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056560.D
 Acq On : 28 Jun 2019 17:13
 Operator : JC/SP
 Sample : K3519-11RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-JMW0882RE

Quant Time: Jun 28 23:32:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYLM
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	47592	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	81020	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	54892	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	12658	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	29053	4.98	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	99.60%
5) Dibromofluoromethane	7.59	113	21527	4.82	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	96.40%
6) Toluene-d8	10.09	98	50669	5.01	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	100.20%
7) 4-Bromofluorobenzene	12.39	95	17908	6.93	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	138.60%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	328	0.069	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056560.D
 Acq On : 28 Jun 2019 17:13
 Operator : JC/SP
 Sample : K3519-11RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-JMW0882RE

Quant Time: Jun 28 23:32:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

