

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062819\
 Data File : VN056546.D
 Acq On : 28 Jun 2019 11:21
 Operator : JC/SP
 Sample : VN0628WBS01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0628WBS01

Quant Time: Jun 28 23:30:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50105	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	85327	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	51088	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	7195	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	31140	5.07	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	101.40%
5) Dibromofluoromethane	7.59	113	23220	4.94	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	98.80%
6) Toluene-d8	10.09	98	50970	4.79	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	95.80%
7) 4-Bromofluorobenzene	12.40	95	12450	4.57	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	91.40%


Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	25685	5.128 ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062819\
Data File : VN056546.D
Acq On : 28 Jun 2019 11:21
Operator : JC/SP
Sample : VN0628WBS01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0628WBS01

Quant Time: Jun 28 23:30:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 5.000 ug/l

RT: 7.65 min Scan# 1243

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21

Instrument :

MSVOA_N

ClientSampleId :


VN0628WBS01

Tot Ion:168 Resp: 50105

Ion Ratio Lower Upper

168 100

99 51.2 41.9 62.9



Abundance Ion 168.00 (167.70 to 168.70): VN056546.D (-1225) (-)

Ion 99.00 (98.70 to 99.70): VN056546.D (-1225) (-)

7.65

15000



10000

5000

0

7.40 7.60 7.80 8.00

Time-->



#2

Vinyl Chloride

Concen: 5.128 ug/l

RT: 2.17 min Scan# 150

Delta R.T. 0.00 min

Lab File: VN056546.D


Acq: 28 Jun 2019 11:21

Tgt Ion: 62 Resp: 25685

Ion Ratio Lower Upper

62 100

64 32.7 26.1 39.1



Abundance Ion 62.00 (61.70 to 62.70): VN056546.D (-98) (-)

Ion 64.00 (63.70 to 64.70): VN056546.D (-98) (-)

2.17

15000


10000

5000

0

2.10 2.20 2.30

Time-->





#3

1,2-Dichloroethane-d4

Concen: 5.074 ug/l

RT: 8.01 min Scan# 1264

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21

Instrument :

MSVOA_N

ClientSampleId :


VN0628WBS01

Tot Ion: 65 Resp: 31140

Ion Ratio Lower Upper

65 100

67 54.1 0.0 108.8



Abundance Ion 65.00 (64.70 to 65.70): VN0

Ion 67.00 (66.70 to 67.70): VN0

8.01

10000

8000

6000



4000

2000

0

Time-->

7.80 8.00 8.20



#4

1,4-Difluorobenzene

Concen: 5.000 ug/l

RT: 8.58 min Scan# 1297

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21


Tgt Ion: 114 Resp: 85327

Ion Ratio Lower Upper

114 100

63 18.7 0.0 37.6

88 15.1 0.0 31.2



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 88.00 (87.70 to 88.70): VN0

8.58

25000

20000

15000


10000

5000

0

Time-->

8.50 9.00





#5

Dibromofluoromethane

Concen: 4.936 ug/l

RT: 7.59 min Scan# 1239

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21

Instrument :

MSVOA_N

ClientSampleId :

VN0628WBS01




Tot Ion:113 Resp: 23220

Ion Ratio Lower Upper


113 100

111 102.4 81.7 122.5

192 20.9 16.9 25.3



Time-->



#6

Toluene-d8

Concen: 4.789 ug/l

RT: 10.09 min Scan# 1440

Delta R.T. 0.00 min

Lab File: VN056546.D




Acq: 28 Jun 2019 11:21

Tgt Ion: 98 Resp: 50970


Ion Ratio Lower Upper

98 100

100 62.1 49.7 74.5



Time-->



#7

4-Bromofluorobenzene

Concen: 4.573 ug/l

RT: 12.40 min Scan# 1773

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21

Instrument :

MSVOA_N

ClientSampleId :

VN0628WBS01





Tot Ion: 95 Resp: 12450

Ion Ratio Lower Upper

95 100

174 85.6 0.0 167.8

176 79.6 0.0 158.0



#8

Chlorobenzene-d5

Concen: 5.000 ug/l

RT: 11.41 min Scan# 1628

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21




Tgt Ion: 117 Resp: 51088



Ion Ratio Lower Upper

117 100

82 53.5 43.0 64.4

119 31.1 24.7 37.1





#9

1,4-Dichlorobenzene-d4

Concen: 5.000 ug/l

RT: 13.35 min Scan# 1864

Delta R.T. 0.00 min

Lab File: VN056546.D

Acq: 28 Jun 2019 11:21

Instrument :

MSVOA_N

ClientSampleId :

VN0628WBS01

Tot Ion:152 Resp: 7195

Ion Ratio Lower Upper

152 100

115 54.4 27.0 80.8

150 152.1 0.0 306.6

Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V

