


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062819\
Data File : VN056539.D
Acq On : 28 Jun 2019 8:19
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Title : SW846 8260
Last Update : Mon Jul 01 15:45:13 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	4476	PASS
75	95	30	60	46.5	13114	PASS
95	95	100	100	100.0	28226	PASS
96	95	5	9	7.4	2096	PASS
173	174	0.00	2	1.1	277	PASS
174	95	50	100	88.8	25061	PASS
175	174	5	9	8.2	2043	PASS
176	174	95	101	95.9	24021	PASS
177	176	5	9	7.0	1670	PASS