

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062821\
 Data File : VN067540.D
 Acq On : 28 Jun 2021 19:04
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Manual Integrations
 APPROVED

MMDadoda
 6/29/2021 6:23:00 PM

Quant Time: Jun 29 05:26:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	227772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	423570	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	399335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	140497	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	165908	53.222	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.440%			
35) Dibromofluoromethane	8.024	113	126958	48.344	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.680%			
50) Toluene-d8	10.446	98	558464	52.350	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.700%			
62) 4-Bromofluorobenzene	12.733	95	177399	47.607	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.220%			
Target Compounds					Qvalue	
59) 2-Hexanone	11.095	43	3871	1.515	ug/l	97
78) n-propylbenzene	12.924	91	3308	0.251	ug/l	97
79) 2-Chlorotoluene	13.010	91	1764	0.219	ug/l	95
82) 4-Chlorotoluene	13.109	91	2692	0.342	ug/l	94
84) 1,2,4-Trimethylbenzene	13.372	105	2817	0.295	ug/l	92
85) sec-Butylbenzene	13.509	105	2860	0.248	ug/l	91
86) p-Isopropyltoluene	13.616	119	2695	0.290	ug/l	91
87) 1,3-Dichlorobenzene	13.613	146	1933	0.398	ug/l	94
88) 1,4-Dichlorobenzene	13.702	146	2354m	0.491	ug/l	
89) n-Butylbenzene	13.948	91	4893	0.639	ug/l	94
91) 1,2-Dichlorobenzene	13.989	146	2087	0.451	ug/l	96
93) 1,2,4-Trichlorobenzene	15.268	180	2967	1.458	ug/l	95
95) Naphthalene	15.509	128	15344	5.257	ug/l	98
96) 1,2,3-Trichlorobenzene	15.700	180	3779	1.897	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062821\
Data File : VN067540.D
Acq On : 28 Jun 2021 19:04
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Manual Integrations
APPROVED

MMDadoda
6/29/2021 6:23:00 PM

Quant Time: Jun 29 05:26:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

