


Naphthalene

Response Ratio



$$\text{Response} = 2.633\text{e}+000 * \text{Amt} - 1.677\text{e}-001$$


Coef of Det (r^2) = 0.998103 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

Acrolein

Response Ratio



$$\text{Response} = 6.790\text{e-}002 * \text{Amt} - 3.024\text{e-}003$$


Coef of Det (r^2) = 0.999427 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

Methylene Chloride

Response Ratio



$$\text{Response} = 5.898\text{e-}001 * \text{Amt} + 2.430\text{e-}003$$


Coef of Det (r^2) = 0.999786 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

Cyclohexane

Response Ratio




$$\text{Response} = 9.024\text{e-}001 * \text{Amt} + 2.952\text{e-}002$$

Coef of Det (r^2) = 0.999848 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021


1,4-Dioxane



Response = $6.665 \times 10^{-3} \times \text{Amt} - 3.532 \times 10^{-3}$
 Coef of Det (r^2) = 0.999545 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M
 Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

t-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 5.943\text{e-}001 * \text{Amt} - 2.681\text{e-}002$$


Coef of Det (r^2) = 0.998950 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

Dibromochloromethane

Response Ratio



$$\text{Response} = 4.423\text{e-}001 * \text{Amt} - 2.134\text{e-}002$$


Coef of Det (r^2) = 0.998997 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

2-Chloroethyl Vinyl ether


Response Ratio



Response = 1.392e-001 * Amt - 6.048e-002
 Coef of Det (r^2) = 0.996415 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M
 Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 1.392\text{e-}001 * \text{Amt} - 6.048\text{e-}002$$


Coef of Det (r^2) = 0.996415 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

Styrene


Response Ratio



Response = 1.302e+000 * Amt - 4.903e-002
Coef of Det (r^2) = 0.999336 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M
Calibration Table Last Updated: Mon Jun 28 14:31:03 2021

Bromoform

Response Ratio



$$\text{Response} = 3.304\text{e-}001 * \text{Amt} - 2.286\text{e-}002$$

Coef of Det (r^2) = 0.998090 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062821W.M

Calibration Table Last Updated: Mon Jun 28 14:31:03 2021