

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
 Data File : VN073210.D
 Acq On : 28 Jun 2022 20:38
 Operator : JC\MD
 Sample : N3494-22DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-3DL

Quant Time: Jun 29 02:03:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	326586	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	529069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	520670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	268946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	198439	42.090	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.180%		
35) Dibromofluoromethane	7.951	113	165304	46.962	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.920%		
50) Toluene-d8	10.374	98	568939	45.509	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.020%		
62) 4-Bromofluorobenzene	12.668	95	273065	56.837	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	113.680%		
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	29259	11.056	ug/l	98
20) Methylene Chloride	5.010	84	62764	13.495	ug/l	91
40) Benzene	8.386	78	51147	3.060	ug/l	97
43) Isopropyl Acetate	8.498	43	23014	1.957	ug/l #	86
65) Chlorobenzene	11.710	112	221922	18.846	ug/l	100
95) Naphthalene	15.433	128	1130190	54.525	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
Data File : VN073210.D
Acq On : 28 Jun 2022 20:38
Operator : JC\MD
Sample : N3494-22DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-3DL

Quant Time: Jun 29 02:03:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

