

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
 Data File : VN073217.D
 Acq On : 28 Jun 2022 23:28
 Operator : JC\MD
 Sample : N3409-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB061722

Quant Time: Jun 29 02:04:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	367052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	604065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	547365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	255515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	228456	43.115	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.220%
35) Dibromofluoromethane	7.951	113	193880	48.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.480%
50) Toluene-d8	10.375	98	639831	44.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.660%
62) 4-Bromofluorobenzene	12.669	95	275347	50.196	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
Data File : VN073217.D
Acq On : 28 Jun 2022 23:28
Operator : JC\MD
Sample : N3409-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB061722

Quant Time: Jun 29 02:04:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

