

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
 Data File : VN073228.D
 Acq On : 29 Jun 2022 03:55
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jun 29 04:28:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	224697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	406688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	369089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	167505	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	186020	57.347	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.700%			
35) Dibromofluoromethane	7.951	113	146064	53.983	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.960%			
50) Toluene-d8	10.374	98	469518	48.858	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.720%			
62) 4-Bromofluorobenzene	12.668	95	186728	50.562	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.120%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.046	85	153605	46.758	ug/l	99
3) Chloromethane	2.269	50	160221	50.236	ug/l	100
4) Vinyl Chloride	2.410	62	137821	50.859	ug/l	100
5) Bromomethane	2.798	94	58268	45.524	ug/l	99
6) Chloroethane	2.963	64	82655	51.851	ug/l	96
7) Trichlorofluoromethane	3.316	101	214653	45.818	ug/l	96
8) Diethyl Ether	3.769	74	103499	54.269	ug/l	80
9) 1,1,2-Trichlorotrifluo...	4.140	101	115457	41.708	ug/l	96
10) Methyl Iodide	4.351	142	147116	44.741	ug/l	94
11) Tert butyl alcohol	5.287	59	212622	270.587	ug/l	99
12) 1,1-Dichloroethene	4.116	96	124181	46.947	ug/l	95
13) Acrolein	3.981	56	129373	260.482	ug/l	97
14) Allyl chloride	4.763	41	276870	49.686	ug/l #	92
15) Acrylonitrile	5.469	53	590051	291.699	ug/l	99
16) Acetone	4.222	43	533320	292.909	ug/l	97
17) Carbon Disulfide	4.463	76	321281	46.842	ug/l	98
18) Methyl Acetate	4.775	43	289939	56.823	ug/l	91
19) Methyl tert-butyl Ether	5.534	73	547142	56.754	ug/l	99
20) Methylene Chloride	5.022	84	165755	51.801	ug/l	92
21) trans-1,2-Dichloroethene	5.516	96	132614	47.981	ug/l	94
22) Diisopropyl ether	6.416	45	587561	56.507	ug/l	96
23) Vinyl Acetate	6.351	43	2687670	288.697	ug/l #	93
24) 1,1-Dichloroethane	6.304	63	290598	54.005	ug/l	99
25) 2-Butanone	7.257	43	844758	291.410	ug/l	90
26) 2,2-Dichloropropane	7.245	77	148406	29.820	ug/l	99
27) cis-1,2-Dichloroethene	7.245	96	169372	50.604	ug/l	98
28) Bromochloromethane	7.581	49	135465	60.111	ug/l	84
29) Tetrahydrofuran	7.604	42	570719	287.528	ug/l	87
30) Chloroform	7.745	83	286733	52.906	ug/l	100
31) Cyclohexane	8.022	56	216466	40.747	ug/l	92
32) 1,1,1-Trichloroethane	7.939	97	233086	48.260	ug/l	99
36) 1,1-Dichloropropene	8.151	75	182347	43.798	ug/l	98
37) Ethyl Acetate	7.333	43	343985	54.681	ug/l	95
38) Carbon Tetrachloride	8.133	117	193093	44.849	ug/l	98
39) Methylcyclohexane	9.392	83	208377	40.202	ug/l	92
40) Benzene	8.386	78	620628	48.303	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
 Data File : VN073228.D
 Acq On : 29 Jun 2022 03:55
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022

Quant Time: Jun 29 04:28:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.563	41	175428	57.501	ug/l #	81
42) 1,2-Dichloroethane	8.457	62	244355	52.464	ug/l	100
43) Isopropyl Acetate	8.492	43	499795	55.282	ug/l	93
44) Trichloroethene	9.151	130	148644	45.460	ug/l	96
45) 1,2-Dichloropropane	9.422	63	168546	51.466	ug/l	99
46) Dibromomethane	9.510	93	115511	51.947	ug/l	93
47) Bromodichloromethane	9.698	83	225981	52.472	ug/l	98
48) Methyl methacrylate	9.492	41	224401	53.607	ug/l	90
49) 1,4-Dioxane	9.504	88	82425	1037.229	ug/l #	92
51) 4-Methyl-2-Pentanone	10.269	43	1627045	283.789	ug/l	90
52) Toluene	10.439	92	367903	46.332	ug/l	100
53) t-1,3-Dichloropropene	10.657	75	231343	48.795	ug/l	98
54) cis-1,3-Dichloropropene	10.127	75	247966	48.445	ug/l #	88
55) 1,1,2-Trichloroethane	10.833	97	166990	51.083	ug/l	99
56) Ethyl methacrylate	10.698	69	293994	55.642	ug/l #	85
57) 1,3-Dichloropropane	10.980	76	287933	51.381	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.980	63	726367	308.248	ug/l	94
59) 2-Hexanone	11.021	43	1236186	282.859	ug/l	89
60) Dibromochloromethane	11.174	129	170022	54.552	ug/l	100
61) 1,2-Dibromoethane	11.280	107	172333	52.779	ug/l	99
64) Tetrachloroethene	10.910	164	135932	45.102	ug/l	97
65) Chlorobenzene	11.710	112	388416	46.531	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.780	131	149927	48.831	ug/l	98
67) Ethyl Benzene	11.780	91	674080	44.481	ug/l	99
68) m/p-Xylenes	11.892	106	515404	90.117	ug/l	96
69) o-Xylene	12.216	106	261051	44.711	ug/l	96
70) Styrene	12.233	104	432507	47.980	ug/l	98
71) Bromoform	12.398	173	122064	54.332	ug/l #	99
73) Isopropylbenzene	12.515	105	655673	45.329	ug/l	99
74) N-amyl acetate	12.327	43	399403	55.144	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.763	83	248000	51.667	ug/l	100
76) 1,2,3-Trichloropropane	12.815	75	217243m	49.600	ug/l	
77) Bromobenzene	12.798	156	158931	47.258	ug/l	85
78) n-propylbenzene	12.857	91	739051	45.992	ug/l	97
79) 2-Chlorotoluene	12.945	91	464355	45.658	ug/l	97
80) 1,3,5-Trimethylbenzene	12.998	105	540925	45.880	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.563	75	77075	48.260	ug/l	96
82) 4-Chlorotoluene	13.039	91	448980	45.600	ug/l	97
83) tert-Butylbenzene	13.262	119	474418	44.884	ug/l	96
84) 1,2,4-Trimethylbenzene	13.304	105	544512	46.259	ug/l	99
85) sec-Butylbenzene	13.439	105	657015	46.291	ug/l	98
86) p-Isopropyltoluene	13.551	119	535006	46.663	ug/l	98
87) 1,3-Dichlorobenzene	13.551	146	275932	45.735	ug/l	99
88) 1,4-Dichlorobenzene	13.627	146	275622	45.595	ug/l	99
89) n-Butylbenzene	13.880	91	438271	45.873	ug/l	98
90) Hexachloroethane	14.145	117	98982	46.324	ug/l	89
91) 1,2-Dichlorobenzene	13.921	146	291917	47.497	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.539	75	65105	55.805	ug/l	89
93) 1,2,4-Trichlorobenzene	15.192	180	150852	47.139	ug/l	97
94) Hexachlorobutadiene	15.298	225	62853	44.182	ug/l	97
95) Naphthalene	15.433	128	649516	50.312	ug/l	100
96) 1,2,3-Trichlorobenzene	15.621	180	155863	47.215	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
Data File : VN073228.D
Acq On : 29 Jun 2022 03:55
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022

Quant Time: Jun 29 04:28:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 1042
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 22469
Ion Ratio Lower Upper
168 100
99 58.0 42.3 63.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#2
Dichlorodifluoromethane
Concen: 46.758 ug/l
RT: 2.046 min Scan# 27
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 85 Resp: 153605
Ion Ratio Lower Upper
85 100
87 31.8 15.7 47.1





#3
Chloromethane
Concen: 50.236 ug/l
RT: 2.269 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 16022
Ion Ratio Lower Upper
50 100
52 32.1 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#4
Vinyl Chloride
Concen: 50.859 ug/l
RT: 2.410 min Scan# 89
Delta R.T. 0.005 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 62 Resp: 137821
Ion Ratio Lower Upper
62 100
64 31.2 25.0 37.4





#5
Bromomethane
Concen: 45.524 ug/l
RT: 2.798 min Scan# 154
Delta R.T. 0.005 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 58268
Ion Ratio Lower Upper
94 100
96 96.4 76.0 114.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#6
Chloroethane
Concen: 51.851 ug/l
RT: 2.963 min Scan# 183
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion: 64 Resp: 82655
Ion Ratio Lower Upper
64 100
66 30.5 26.2 39.4









#7
Trichlorofluoromethane
Concen: 45.818 ug/l
RT: 3.316 min Scan# 243
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 21465
Ion Ratio Lower Upper
101 100
103 61.6 51.5 77.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#8
Diethyl Ether
Concen: 54.269 ug/l
RT: 3.769 min Scan# 320
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion: 74 Resp: 103499
Ion Ratio Lower Upper
74 100
45 104.4 42.9 128.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 41.708 ug/l
RT: 4.140 min Scan# 383
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 11545
Ion Ratio Lower Upper
101 100
85 44.6 34.7 52.1
151 74.6 63.0 94.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#10
Methyl Iodide
Concen: 44.741 ug/l
RT: 4.351 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion:142 Resp: 147116
Ion Ratio Lower Upper
142 100
127 43.7 31.3 46.9
141 14.6 11.2 16.8





#11
Tert butyl alcohol
Concen: 270.587 ug/l
RT: 5.287 min Scan# 51
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 59 Resp: 21262
Ion Ratio Lower Upper
59 100
57 10.6 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#12
1,1-Dichloroethene
Concen: 46.947 ug/l
RT: 4.116 min Scan# 379
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 96 Resp: 124181
Ion Ratio Lower Upper
96 100
61 168.2 127.6 191.4
98 65.6 51.2 76.8





#13
Acrolein
Concen: 260.482 ug/l
RT: 3.981 min Scan# 356
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 12937
Ion Ratio Lower Upper
56 100
55 71.2 54.8 82.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#14
Allyl chloride
Concen: 49.686 ug/l
RT: 4.763 min Scan# 489
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 41 Resp: 276870
Ion Ratio Lower Upper
41 100
39 63.9 53.2 79.8
76 28.3 29.9 44.9#





#15
Acrylonitrile
Concen: 291.699 ug/l
RT: 5.469 min Scan# 609
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 59005
Ion Ratio Lower Upper
53 100
52 81.5 66.4 99.6
51 35.5 28.5 42.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#16
Acetone
Concen: 292.909 ug/l
RT: 4.222 min Scan# 397
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion: 43 Resp: 533320
Ion Ratio Lower Upper
43 100
58 30.6 26.0 39.0









#17
Carbon Disulfide
Concen: 46.842 ug/l
RT: 4.463 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 76 Resp: 32128
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#18
Methyl Acetate
Concen: 56.823 ug/l
RT: 4.775 min Scan# 491
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 43 Resp: 289939
Ion Ratio Lower Upper
43 100
74 22.1 21.6 32.4





#19
Methyl tert-butyl Ether
Concen: 56.754 ug/l
RT: 5.534 min Scan# 619
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 73 Resp: 54714
Ion Ratio Lower Upper
73 100
57 21.7 17.6 26.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#20
Methylene Chloride
Concen: 51.801 ug/l
RT: 5.022 min Scan# 533
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 84 Resp: 165755
Ion Ratio Lower Upper
84 100
49 120.8 88.3 132.5
51 38.8 28.2 42.2
86 59.8 50.6 76.0





#21
trans-1,2-Dichloroethene
Concen: 47.981 ug/l
RT: 5.516 min Scan# 616
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 132614
Ion Ratio Lower Upper
96 100
61 141.6 105.8 158.8
98 63.7 50.3 75.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#22
Diisopropyl ether
Concen: 56.507 ug/l
RT: 6.416 min Scan# 770
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 45 Resp: 587561
Ion Ratio Lower Upper
45 100
43 58.7 46.6 70.0
87 24.6 24.3 36.5
59 10.9 9.6 14.4





#23
 Vinyl Acetate
 Concen: 288.697 ug/l
 RT: 6.351 min Scan# 759
 Delta R.T. 0.000 min
 Lab File: VN073228.D
 Acq: 29 Jun 2022 03:55




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 2687670
 Ion Ratio Lower Upper
 43 100
 86 9.9 10.0 15.0

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022




#24
 1,1-Dichloroethane
 Concen: 54.005 ug/l
 RT: 6.304 min Scan# 751
 Delta R.T. 0.000 min
 Lab File: VN073228.D
 Acq: 29 Jun 2022 03:55




Tgt Ion: 63 Resp: 290598
 Ion Ratio Lower Upper
 63 100
 98 6.3 3.5 10.4
 100 4.6 2.3 6.8





#25
2-Butanone
Concen: 291.410 ug/l
RT: 7.257 min Scan# 913
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 844758
Ion Ratio Lower Upper
43 100
72 24.3 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#26
2,2-Dichloropropane
Concen: 29.820 ug/l
RT: 7.245 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 77 Resp: 148406
Ion Ratio Lower Upper
77 100
97 23.3 11.4 34.2





#27
cis-1,2-Dichloroethene
Concen: 50.604 ug/l
RT: 7.245 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 169372
Ion Ratio Lower Upper
96 100
61 143.5 0.0 280.6
98 64.8 0.0 128.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#28
Bromochloromethane
Concen: 60.111 ug/l
RT: 7.581 min Scan# 968
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion: 49 Resp: 135465
Ion Ratio Lower Upper
49 100
129 2.0 0.0 4.6
130 71.9 69.7 104.5





#29

Tetrahydrofuran

Concen: 287.528 ug/l

RT: 7.604 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 570719

Ion	Ratio	Lower	Upper
42	100		
72	39.9	39.7	59.5
71	37.7	36.6	55.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#30

Chloroform

Concen: 52.906 ug/l

RT: 7.745 min Scan# 996




Delta R.T. 0.000 min


Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Tgt Ion: 83 Resp: 286733


Ion	Ratio	Lower	Upper
83	100		
85	64.5	51.8	77.8





#31
Cyclohexane
Concen: 40.747 ug/l
RT: 8.022 min Scan# 1043
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 216460
Ion Ratio Lower Upper
56 100
69 33.0 27.8 41.8
84 86.9 76.9 115.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#32
1,1,1-Trichloroethane
Concen: 48.260 ug/l
RT: 7.939 min Scan# 1029
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55



Tgt Ion: 97 Resp: 233086
Ion Ratio Lower Upper
97 100
99 63.9 50.6 76.0
61 46.3 36.0 54.0





#33

1,2-Dichloroethane-d4

Concen: 57.347 ug/l

RT: 8.369 min Scan# 1102

Delta R.T. -0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 186020

Ion Ratio Lower Upper

65 100





67 53.2 0.0 104.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022

Supervised By :Semsettin Yesilyurt 06/30/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.904 min Scan# 1193

Delta R.T. 0.006 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55




Tgt Ion:114 Resp: 406688


Ion Ratio Lower Upper

114 100

63 20.1 0.0 40.0

88 15.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 53.983 ug/l

RT: 7.951 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VN073228.D


Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	113	Resp:	146064
Ion Ratio	Lower	Upper	
113	100		
111	102.6	81.4	122.0
192	18.9	17.0	25.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#36

1,1-Dichloropropene

Concen: 43.798 ug/l




RT: 8.151 min Scan# 1065


Delta R.T. 0.006 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55


Tgt Ion:	75	Resp:	182347
Ion Ratio	Lower	Upper	
75	100		
110	35.0	18.3	54.8
77	30.0	25.1	37.7





#37
Ethyl Acetate
Concen: 54.681 ug/l
RT: 7.333 min Scan# 91
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 34398
Ion Ratio Lower Upper
43 100
61 13.0 11.8 17.6
70 9.6 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#38
Carbon Tetrachloride
Concen: 44.849 ug/l
RT: 8.133 min Scan# 1062
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion:117 Resp: 193093
Ion Ratio Lower Upper
117 100
119 94.4 75.8 113.8
121 27.7 24.6 37.0






#39
Methylcyclohexane
Concen: 40.202 ug/l
RT: 9.392 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 83 Resp: 20837
Ion Ratio Lower Upper
83 100
55 82.5 57.8 86.8
98 50.8 39.4 59.2


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#40
Benzene
Concen: 48.303 ug/l
RT: 8.386 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 78 Resp: 620628
Ion Ratio Lower Upper
78 100
77 24.4 19.4 29.0





#41
Methacrylonitrile
Concen: 57.501 ug/l
RT: 7.563 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 41 Resp: 175428
Ion Ratio Lower Upper
41 100
39 53.4 48.6 73.0
67 66.0 73.4 110.2
52 27.4 27.7 41.5#

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#42
1,2-Dichloroethane
Concen: 52.464 ug/l
RT: 8.457 min Scan# 1117
Delta R.T. -0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 62 Resp: 244355
Ion Ratio Lower Upper
62 100
98 9.6 0.0 19.4





#43
Isopropyl Acetate
Concen: 55.282 ug/l
RT: 8.492 min Scan# 1123
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 499799
Ion Ratio Lower Upper
43 100
61 23.6 21.9 32.9
87 12.2 12.0 18.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#44
Trichloroethene
Concen: 45.460 ug/l
RT: 9.151 min Scan# 1235
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion:130 Resp: 148644
Ion Ratio Lower Upper
130 100
95 97.2 0.0 186.8





#45
1,2-Dichloropropane
Concen: 51.466 ug/l
RT: 9.422 min Scan# 1281
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 63 Resp: 168540
Ion Ratio Lower Upper
63 100
65 31.9 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#46
Dibromomethane
Concen: 51.947 ug/l
RT: 9.510 min Scan# 1296
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 93 Resp: 115511
Ion Ratio Lower Upper
93 100
95 82.3 66.9 100.3
174 96.1 86.6 130.0





#47
Bromodichloromethane
Concen: 52.472 ug/l
RT: 9.698 min Scan# 1328
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 22598
Ion Ratio Lower Upper
83 100
85 64.4 50.5 75.7
127 8.7 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#48
Methyl methacrylate
Concen: 53.607 ug/l
RT: 9.492 min Scan# 1293
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion: 41 Resp: 224401
Ion Ratio Lower Upper
41 100
69 82.1 74.7 112.1
39 62.2 45.6 68.4





#49

1,4-Dioxane

Concen: 1037.229 ug/l

RT: 9.504 min Scan# 1295

Delta R.T. -0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 88 Resp: 82425
 Ion Ratio Lower Upper
 88 100
 43 33.4 20.2 30.4
 58 71.9 54.2 81.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022



#50

Toluene-d8

Concen: 48.858 ug/l




RT: 10.374 min Scan# 1443


Delta R.T. -0.001 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Tgt Ion: 98 Resp: 469518
 Ion Ratio Lower Upper
 98 100
 100 66.2 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 283.789 ug/l
RT: 10.269 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 162704
Ion Ratio Lower Upper
43 100
58 36.4 34.1 51.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#52
Toluene
Concen: 46.332 ug/l
RT: 10.439 min Scan# 1454
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 92 Resp: 367903
Ion Ratio Lower Upper
92 100
91 172.9 138.6 207.8





#53
t-1,3-Dichloropropene
Concen: 48.795 ug/l
RT: 10.657 min Scan# 1491
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 23134
Ion Ratio Lower Upper
75 100
77 33.0 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#54
cis-1,3-Dichloropropene
Concen: 48.445 ug/l
RT: 10.127 min Scan# 1401
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 75 Resp: 247966
Ion Ratio Lower Upper
75 100
77 33.2 25.2 37.8
39 56.6 35.7 53.5#





#55
1,1,2-Trichloroethane
Concen: 51.083 ug/l
RT: 10.833 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 97 Resp: 166990
Ion Ratio Lower Upper
97 100
83 84.4 66.2 99.4
85 56.9 43.9 65.9
99 63.4 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#56
Ethyl methacrylate
Concen: 55.642 ug/l
RT: 10.698 min Scan# 1498
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55



Tgt Ion: 69 Resp: 293994
Ion Ratio Lower Upper
69 100
41 74.3 47.8 71.6#
39 37.8 27.3 40.9





#57

1,3-Dichloropropane

Concen: 51.381 ug/l

RT: 10.980 min Scan# 1546

Delta R.T. 0.000 min

Lab File: VN073228.D


Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 76 Resp: 28793

Ion Ratio Lower Upper




76 100

78 32.1 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#58

2-Chloroethyl Vinyl ether

Concen: 308.248 ug/l

RT: 9.980 min Scan# 1376

Delta R.T. 0.005 min

Lab File: VN073228.D




Acq: 29 Jun 2022 03:55


Tgt Ion: 63 Resp: 726367

Ion Ratio Lower Upper

63 100


106 27.5 24.6 36.8





#59
2-Hexanone
Concen: 282.859 ug/l
RT: 11.021 min Scan# 1553
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 123618
Ion Ratio Lower Upper
43 100
58 51.3 29.7 89.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#60
Dibromochloromethane
Concen: 54.552 ug/l
RT: 11.174 min Scan# 1579
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion:129 Resp: 170022
Ion Ratio Lower Upper
129 100
127 77.2 38.8 116.4





#61
1,2-Dibromoethane
Concen: 52.779 ug/l
RT: 11.280 min Scan# 1597
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 17233
Ion Ratio Lower Upper
107 100
109 94.5 75.0 112.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#62
4-Bromofluorobenzene
Concen: 50.562 ug/l
RT: 12.668 min Scan# 1833
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 95 Resp: 186728
Ion Ratio Lower Upper
95 100
174 76.1 0.0 167.6
176 70.5 0.0 163.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.680 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:117 Resp: 369089
Ion Ratio Lower Upper
117 100
82 56.3 44.2 66.2
119 32.7 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#64
Tetrachloroethene
Concen: 45.102 ug/l
RT: 10.910 min Scan# 1534
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion:164 Resp: 135932
Ion Ratio Lower Upper
164 100
166 123.8 104.2 156.2
129 92.3 73.9 110.9
131 90.3 69.4 104.2





#65
Chlorobenzene
Concen: 46.531 ug/l
RT: 11.710 min Scan# 1669
Delta R.T. 0.006 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 388410
Ion Ratio Lower Upper
112 100
114 30.0 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 48.831 ug/l
RT: 11.780 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55



Tgt Ion:131 Resp: 149927
Ion Ratio Lower Upper
131 100
133 94.9 47.8 143.4
119 64.4 33.4 100.2









#67
Ethyl Benzene
Concen: 44.481 ug/l
RT: 11.780 min Scan# 1107
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 674080
Ion Ratio Lower Upper
91 100
106 30.8 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#68
m/p-Xylenes
Concen: 90.117 ug/l
RT: 11.892 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion:106 Resp: 515404
Ion Ratio Lower Upper
106 100
91 205.0 158.9 238.3









#69
o-Xylene
Concen: 44.711 ug/l
RT: 12.216 min Scan# 1756
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 26105
Ion Ratio Lower Upper
106 100
91 218.6 106.0 318.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#70
Styrene
Concen: 47.980 ug/l
RT: 12.233 min Scan# 1759
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55

Tgt Ion:104 Resp: 432507
Ion Ratio Lower Upper
104 100
78 51.8 39.5 59.3
103 55.0 44.0 66.0





#71

Bromoform

Concen: 54.332 ug/l

RT: 12.398 min Scan# 1787

Delta R.T. -0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:173	Resp: 122064
Ion Ratio	Lower Upper
173 100	
175 49.9	24.6 73.6
254 0.0	0.1 0.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.610 min Scan# 1993


Delta R.T. -0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55


Tgt Ion:152	Resp: 167505
Ion Ratio	Lower Upper
152 100	
115 61.1	28.9 86.7
150 174.2	0.0 356.2





#73
Isopropylbenzene
Concen: 45.329 ug/l
RT: 12.515 min Scan# 1807
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 655673
Ion Ratio Lower Upper
105 100
120 26.4 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#74
N-ethyl acetate
Concen: 55.144 ug/l
RT: 12.327 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 43 Resp: 399403
Ion Ratio Lower Upper
43 100
70 38.5 39.0 58.4#
55 22.5 21.1 31.7
61 22.3 21.1 31.7





#75
1,1,2,2-Tetrachloroethane
Concen: 51.667 ug/l
RT: 12.763 min Scan# 1849
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument : MSVOA_N
ClientSampleId : VSTDCCC050EC




Tgt Ion: 83 Resp: 248000
Ion Ratio Lower Upper
83 100
131 10.0 5.3 15.9
85 64.8 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#76
1,2,3-Trichloropropane
Concen: 49.600 ug/l m
RT: 12.815 min Scan# 1858
Delta R.T. -0.001 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion: 75 Resp: 217243
Ion Ratio Lower Upper
75 100
77 172.0 98.5 295.5





#77
Bromobenzene
Concen: 47.258 ug/l
RT: 12.798 min Scan# 1855
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:156 Resp: 15893
Ion Ratio Lower Upper
156 100
77 235.1 100.5 301.5
158 97.0 48.9 146.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#78
n-propylbenzene
Concen: 45.992 ug/l
RT: 12.857 min Scan# 1865
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55



Tgt Ion: 91 Resp: 739051
Ion Ratio Lower Upper
91 100
120 22.3 11.9 35.5





#79

2-Chlorotoluene

Concen: 45.658 ug/l

RT: 12.945 min Scan# 1880

Delta R.T. -0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 46435

Ion Ratio Lower Upper

91 100





126 32.6 17.3 51.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022

Supervised By :Semsettin Yesilyurt 06/30/2022



#80

1,3,5-Trimethylbenzene

Concen: 45.880 ug/l

RT: 12.998 min Scan# 1889

Delta R.T. -0.000 min

Lab File: VN073228.D




Acq: 29 Jun 2022 03:55


Tgt Ion:105 Resp: 540925

Ion Ratio Lower Upper

105 100


120 48.1 24.9 74.7





#81
trans-1,4-Dichloro-2-butene
Concen: 48.260 ug/l
RT: 12.563 min Scan# 1815
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 77075
Ion Ratio Lower Upper
75 100
53 103.6 78.0 117.0
89 47.0 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#82
4-Chlorotoluene
Concen: 45.600 ug/l
RT: 13.039 min Scan# 1896
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion: 91 Resp: 448980
Ion Ratio Lower Upper
91 100
126 32.4 17.2 51.6





#83
tert-Butylbenzene
Concen: 44.884 ug/l
RT: 13.262 min Scan# 1934
Delta R.T. 0.005 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 474418
Ion Ratio Lower Upper
119 100
91 67.0 31.6 94.8
134 26.8 13.7 41.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#84
1,2,4-Trimethylbenzene
Concen: 46.259 ug/l
RT: 13.304 min Scan# 1941
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55



Tgt Ion:105 Resp: 544512
Ion Ratio Lower Upper
105 100
120 45.6 23.3 69.8









#85
 sec-Butylbenzene
 Concen: 46.291 ug/l
 RT: 13.439 min Scan# 1964
 Delta R.T. -0.000 min
 Lab File: VN073228.D
 Acq: 29 Jun 2022 03:55

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion:105 Resp: 657015
 Ion Ratio Lower Upper
 105 100
 134 20.0 10.4 31.1




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022



#86
 p-Isopropyltoluene
 Concen: 46.663 ug/l
 RT: 13.551 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN073228.D
 Acq: 29 Jun 2022 03:55


Tgt Ion:119 Resp: 535006
 Ion Ratio Lower Upper
 119 100
 134 26.0 13.3 39.8
 91 24.0 11.3 33.9





#87
1,3-Dichlorobenzene
Concen: 45.735 ug/l
RT: 13.551 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 275932
Ion Ratio Lower Upper
146 100
111 40.2 20.0 60.0
148 62.6 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#88
1,4-Dichlorobenzene
Concen: 45.595 ug/l
RT: 13.627 min Scan# 1996
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Tgt Ion:146 Resp: 275622
Ion Ratio Lower Upper
146 100
111 38.9 19.1 57.1
148 62.8 31.4 94.3





#89
n-Butylbenzene
Concen: 45.873 ug/l
RT: 13.880 min Scan# 2039
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 43827
Ion Ratio Lower Upper
91 100
92 54.3 26.9 80.7
134 25.4 13.6 40.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022



#90
Hexachloroethane
Concen: 46.324 ug/l
RT: 14.145 min Scan# 2084
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55


Tgt Ion:117 Resp: 98982
Ion Ratio Lower Upper
117 100
201 71.1 40.6 122.0





#91
1,2-Dichlorobenzene
Concen: 47.497 ug/l
RT: 13.921 min Scan# 2046
Delta R.T. 0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 291917
Ion Ratio Lower Upper
146 100
111 41.8 20.9 62.8
148 62.4 32.1 96.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 55.805 ug/l
RT: 14.539 min Scan# 2151
Delta R.T. -0.000 min
Lab File: VN073228.D
Acq: 29 Jun 2022 03:55



Tgt Ion: 75 Resp: 65105
Ion Ratio Lower Upper
75 100
155 82.1 44.9 134.5
157 103.5 59.0 176.8





#93

1,2,4-Trichlorobenzene

Concen: 47.139 ug/l

RT: 15.192 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN073228.D


Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion:180 Resp: 150852

Ion Ratio Lower Upper

180 100

182 97.2 47.5 142.5




145 31.6 14.5 43.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022

Supervised By :Semsettin Yesilyurt 06/30/2022



#94

Hexachlorobutadiene

Concen: 44.182 ug/l

RT: 15.298 min Scan# 2280

Delta R.T. -0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55




Tgt Ion:225 Resp: 62853


Ion Ratio Lower Upper

225 100

223 62.2 32.4 97.2

227 61.4 31.8 95.4





#95

Naphthalene

Concen: 50.312 ug/l

RT: 15.433 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:128 Resp: 649510
 Ion Ratio Lower Upper
 128 100
 127 12.7 10.3 15.5
 129 10.6 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022



#96

1,2,3-Trichlorobenzene

Concen: 47.215 ug/l

RT: 15.621 min Scan# 2335

Delta R.T. 0.000 min

Lab File: VN073228.D

Acq: 29 Jun 2022 03:55

Tgt Ion:180 Resp: 155863
 Ion Ratio Lower Upper
 180 100
 182 97.8 47.8 143.3
 145 33.5 15.5 46.5

