



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062822\
Data File : VN073204.D
Acq On : 28 Jun 2022 17:38
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Title : SW846 8260
Last Update : Wed Jun 29 01:54:03 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	12845	PASS
75	95	30	60	54.8	34269	PASS
95	95	100	100	100.0	62520	PASS
96	95	5	9	6.4	3978	PASS
173	174	0.00	2	1.5	724	PASS
174	95	50	100	76.8	48035	PASS
175	174	5	9	7.5	3593	PASS
176	174	95	101	95.4	45845	PASS
177	176	5	9	7.4	3396	PASS