

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
 Data File : VN078433.D
 Acq On : 28 Jun 2023 11:39
 Operator : JC\MD
 Sample : VN0628WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0628WBL01

Quant Time: Jun 29 05:17:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	463799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	831662	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	746924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	229544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	313358	50.674	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.340%	
35) Dibromofluoromethane	8.177	113	297772	56.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 113.260%	
50) Toluene-d8	10.571	98	967481	47.725	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.440%	
62) 4-Bromofluorobenzene	12.853	95	306599	48.651	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
Data File : VN078433.D
Acq On : 28 Jun 2023 11:39
Operator : JC\MD
Sample : VN0628WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0628WBL01

Quant Time: Jun 29 05:17:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

