

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
 Data File : VN078443.D
 Acq On : 28 Jun 2023 15:54
 Operator : JC\MD
 Sample : 03310-06RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-EB-20230613RE

Quant Time: Jun 29 05:22:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	443629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	826200	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	740063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	224300	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	309071	52.253	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.500%
35) Dibromofluoromethane	8.177	113	292711	56.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.060%
50) Toluene-d8	10.570	98	953939	47.368	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.740%
62) 4-Bromofluorobenzene	12.853	95	299152	47.783	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
Data File : VN078443.D
Acq On : 28 Jun 2023 15:54
Operator : JC\MD
Sample : 03310-06RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-EB-20230613RE

Quant Time: Jun 29 05:22:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

