

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
 Data File : VN078444.D
 Acq On : 28 Jun 2023 16:18
 Operator : JC\MD
 Sample : 03310-03RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-883-885RE

Quant Time: Jun 29 05:22:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/29/2023
 Supervised By : Mahesh Dadoda 06/29/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	444301	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	812464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	733837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	226023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	306199	51.689	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.380%
35) Dibromofluoromethane	8.171	113	294547	57.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.680%
50) Toluene-d8	10.571	98	934244	47.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.340%
62) 4-Bromofluorobenzene	12.853	95	298322	48.456	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%
Target Compounds						
16) Acetone	4.442	43	4521m	2.651	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
Data File : VN078444.D
Acq On : 28 Jun 2023 16:18
Operator : JC\MD
Sample : 03310-03RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-883-885RE

Quant Time: Jun 29 05:22:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/29/2023
Supervised By :Mahesh Dadoda 06/29/2023

