

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
 Data File : VN078446.D
 Acq On : 28 Jun 2023 17:07
 Operator : JC\MD
 Sample : 03310-05RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-943-945RE

Quant Time: Jun 29 05:23:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	437442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	806072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	710139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	221476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	303157	51.978	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.960%
35) Dibromofluoromethane	8.177	113	287712	56.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.900%
50) Toluene-d8	10.571	98	912612	46.448	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.900%
62) 4-Bromofluorobenzene	12.859	95	293048	47.977	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.960%
Target Compounds						
16) Acetone	4.442	43	35918	21.395	ug/l #	86
25) 2-Butanone	7.500	43	23602	9.142	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
Data File : VN078446.D
Acq On : 28 Jun 2023 17:07
Operator : JC\MD
Sample : 03310-05RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-943-945RE

Quant Time: Jun 29 05:23:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

