


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062823\
Data File : VN078430.D
Acq On : 28 Jun 2023 09:33
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Title : SW846 8260
Last Update : Wed Jun 07 05:05:00 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	14856	PASS
75	95	30	60	52.7	40253	PASS
95	95	100	100	100.0	76336	PASS
96	95	5	9	7.2	5462	PASS
173	174	0.00	2	1.3	747	PASS
174	95	50	100	75.5	57653	PASS
175	174	5	9	8.1	4664	PASS
176	174	95	101	97.6	56280	PASS
177	176	5	9	6.6	3704	PASS

Average of 12.847 to 12.859 min.: VN078430.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.90	479	47.60	108	60.90	2437	70.20	191
36.20	158	47.95	562	61.10	1509	71.70	121
37.00	4596	48.95	3318	62.00	4073	71.95	221
37.95	3465	50.05	14856	62.95	3522	72.95	393
38.95	1332	51.00	4605	64.00	686	74.05	14768
43.00	60	52.20	57	65.05	622	75.00	40253
44.00	676	52.80	50	65.70	53	76.05	3462
44.95	837	54.95	199	67.05	221	76.90	584
46.00	68	56.00	1562	68.00	7674	77.70	101
46.90	579	57.00	2143	69.00	8654	77.90	295
47.10	273	59.90	791	69.90	623	78.10	101

Average of 12.847 to 12.859 min.: VN078430.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
78.85	3486	94.00	10768	115.60	78	134.90	70
79.90	863	95.00	76336	115.95	248	136.95	113
80.95	4266	96.05	5462	116.90	547	140.95	954
81.85	595	97.10	236	117.60	84	142.95	1086
86.20	220	103.90	485	117.90	354	143.70	73
86.95	3133	105.10	55	118.85	784	144.90	104
87.90	3867	105.60	101	127.85	291	145.95	173
90.90	239	105.90	243	128.70	215	146.95	453
91.10	59	110.10	55	129.75	192	148.10	306
91.95	2320	110.90	55	130.85	174	149.85	161
92.95	4022	114.70	53	134.40	71	152.80	66

Average of 12.847 to 12.859 min.: VN078430.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
155.10	180	175.95	56280				
156.80	73	176.95	3704				
159.00	72	204.90	50				
170.50	56	207.90	57				
170.75	143	281.05	109				
171.80	209						
171.95	698						
172.80	241						
173.05	747						
173.90	57653						
175.00	4664						