

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062918\
 Data File : VN049639.D
 Acq On : 29 Jun 2018 18:41
 Operator : MD\SY
 Sample : J3612-03RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-3(5-5.5)RE

Quant Time: Jun 30 01:54:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 29 15:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1169896	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1874526	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1577619	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	557981	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	762304	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
35) Dibromofluoromethane	7.60	113	23732	1.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	3.38%	
50) Toluene-d8	10.09	98	2547118	46.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.86%	
62) 4-Bromofluorobenzene	12.40	95	717077	38.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.58%	
Target Compounds						
16) Acetone	3.83	43	141712	41.67	ug/l #	89
20) Methylene Chloride	4.55	84	17121	1.08	ug/l #	82
25) 2-Butanone	6.85	43	14771	2.88	ug/l	97
64) Tetrachloroethene	10.63	164	972106	45.51	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062918\
Data File : VN049639.D
Acq On : 29 Jun 2018 18:41
Operator : MD\SY
Sample : J3612-03RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-3(5-5.5)RE

Quant Time: Jun 30 01:54:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 29 15:11:52 2018
Response via : Initial Calibration

