

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062918\
 Data File : VN049640.D
 Acq On : 29 Jun 2018 19:06
 Operator : MD\SY
 Sample : J3612-04RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-3(8.5-9)RE

Quant Time: Jun 30 01:59:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 29 15:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1061821	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1671846	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1381871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	506519	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	688321	49.09	ug/l	0.00
Spiked Amount			Recovery	=	98.18%	
35) Dibromofluoromethane	7.60	113	23597	1.89	ug/l	0.00
Spiked Amount			Recovery	=	3.78%	
50) Toluene-d8	10.09	98	2250256	46.00	ug/l	0.00
Spiked Amount			Recovery	=	92.00%	
62) 4-Bromofluorobenzene	12.40	95	629331	37.68	ug/l	0.00
Spiked Amount			Recovery	=	75.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	112464	36.19	ug/l	93
20) Methylene Chloride	4.54	84	18828	1.30	ug/l	# 83
25) 2-Butanone	6.84	43	12839	2.76	ug/l	95
64) Tetrachloroethene	10.63	164	588437	31.45	ug/l	97
95) Naphthalene	15.14	128	17131	3.49	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062918\
Data File : VN049640.D
Acq On : 29 Jun 2018 19:06
Operator : MD\SY
Sample : J3612-04RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-3(8.5-9)RE

Quant Time: Jun 30 01:59:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 29 15:11:52 2018
Response via : Initial Calibration

