

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062918\
 Data File : VN049641.D
 Acq On : 29 Jun 2018 19:32
 Operator : MD\SY
 Sample : J3612-05RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-4(3-3.5)RE

Quant Time: Jun 30 02:02:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 29 15:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1105194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1786035	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1513963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	557590	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	749729	51.37	ug/l	0.00
Spiked Amount			Recovery	=	102.74%	
35) Dibromofluoromethane	7.59	113	31392	2.35	ug/l	0.00
Spiked Amount			Recovery	=	4.70%	
50) Toluene-d8	10.09	98	2442716	46.74	ug/l	0.00
Spiked Amount			Recovery	=	93.48%	
62) 4-Bromofluorobenzene	12.40	95	695943	39.00	ug/l	0.00
Spiked Amount			Recovery	=	78.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	59083	17.35	ug/l	94
20) Methylene Chloride	4.55	84	21411	1.42	ug/l	88
25) 2-Butanone	6.85	43	9228	1.90	ug/l	95
64) Tetrachloroethene	10.63	164	45389	2.21	ug/l	96
95) Naphthalene	15.14	128	8122	3.00	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062918\
Data File : VN049641.D
Acq On : 29 Jun 2018 19:32
Operator : MD\SY
Sample : J3612-05RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-4(3-3.5)RE

Quant Time: Jun 30 02:02:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 29 15:11:52 2018
Response via : Initial Calibration

