

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN062920\
 Data File : VN062182.D
 Acq On : 29 Jun 2020 11:40
 Operator : JC/MD
 Sample : VN0629MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0629MBL01

Quant Time: Jun 30 04:49:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	208278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	390008	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	358412	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	123930	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	151375	51.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.58%	
35) Dibromofluoromethane	7.54	113	119733	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	
50) Toluene-d8	10.06	98	486986	48.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.38%	
62) 4-Bromofluorobenzene	12.38	95	151822	44.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062182.D
Acq On : 29 Jun 2020 11:40
Operator : JC/MD
Sample : VN0629MBL01
Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0629MBL01

Quant Time: Jun 30 04:49:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

