

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062188.D
 Acq On : 29 Jun 2020 15:17
 Operator : JC/MD
 Sample : L3088-01DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 08BR-20200622DL

Quant Time: Jun 30 05:04:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	212165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	405811	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	389317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	146375	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	155096	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
35) Dibromofluoromethane	7.55	113	127644	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	10.06	98	515539	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.38	95	169630	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	1790	0.735	ug/l #	77
27) cis-1,2-Dichloroethene	6.79	96	8963	2.870	ug/l	88
44) Trichloroethene	8.80	130	152087	48.494	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062188.D
Acq On : 29 Jun 2020 15:17
Operator : JC/MD
Sample : L3088-01DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
08BR-20200622DL

Quant Time: Jun 30 05:04:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

