

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062193.D
 Acq On : 29 Jun 2020 17:19
 Operator : JC/MD
 Sample : L3088-03DL 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20BR-20200622DL

Quant Time: Jun 30 05:16:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	196100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	364443	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	342931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	113094	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	144818	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	7.55	113	117462	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
50) Toluene-d8	10.06	98	463647	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	12.38	95	139500	44.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.36%	
Target Compounds						
4) Vinyl Chloride	2.16	62	26294	7.533	ug/l	94
27) cis-1,2-Dichloroethene	6.78	96	147117	50.968	ug/l	91
44) Trichloroethene	8.80	130	27967	9.930	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062193.D
Acq On : 29 Jun 2020 17:19
Operator : JC/MD
Sample : L3088-03DL 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20BR-20200622DL

Quant Time: Jun 30 05:16:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

