

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062198.D
 Acq On : 29 Jun 2020 19:20
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 30 05:26:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	191682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	368784	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	353322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125175	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	144698	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
35) Dibromofluoromethane	7.55	113	119777	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
50) Toluene-d8	10.06	98	469532	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
62) 4-Bromofluorobenzene	12.38	95	147655	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
Target Compounds						
44) Trichloroethene	8.80	130	1384	0.486	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062198.D
Acq On : 29 Jun 2020 19:20
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 30 05:26:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

