

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062199.D
 Acq On : 29 Jun 2020 19:44
 Operator : JC/MD
 Sample : L3127-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BRP1-20200625

Quant Time: Jun 30 05:36:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	193276	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	362435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	346144	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	122685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	140001	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	7.55	113	116499	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
50) Toluene-d8	10.06	98	458521	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	12.38	95	143857	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
Target Compounds						
4) Vinyl Chloride	2.16	62	55830	16.228	ug/l	100
12) 1,1-Dichloroethene	3.70	96	1360	0.613	ug/l #	65
27) cis-1,2-Dichloroethene	6.78	96	171440	60.263	ug/l	87
44) Trichloroethene	8.80	130	7648	2.730	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062199.D
Acq On : 29 Jun 2020 19:44
Operator : JC/MD
Sample : L3127-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BRP1-20200625

Quant Time: Jun 30 05:36:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

