

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062210.D
 Acq On : 30 Jun 2020 1:22
 Operator : JC/MD
 Sample : PB129816TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129816TB

Quant Time: Jun 30 06:19:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	207215	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	400081	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	406689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	145891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	159541	54.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.72%	
35) Dibromofluoromethane	7.55	113	128401	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
50) Toluene-d8	10.06	98	536036	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
62) 4-Bromofluorobenzene	12.38	95	172415	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
Target Compounds						
16) Acetone	3.78	43	9321	6.262	ug/l	98
20) Methylene Chloride	4.50	84	9697	3.431	ug/l	96
43) Isopropyl Acetate	8.13	43	53574	9.869	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062210.D
Acq On : 30 Jun 2020 1:22
Operator : JC/MD
Sample : PB129816TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129816TB

Quant Time: Jun 30 06:19:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

