

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062221.D
 Acq On : 30 Jun 2020 5:48
 Operator : JC/MD
 Sample : L2806-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-06-062620

Quant Time: Jun 30 06:50:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	207213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	412061	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	411967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	144405	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	160041	55.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.08%	
35) Dibromofluoromethane	7.55	113	131732	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
50) Toluene-d8	10.06	98	543309	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
62) 4-Bromofluorobenzene	12.38	95	170614	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
Target Compounds						
16) Acetone	3.78	43	30060	29.358	ug/l	96
18) Methyl Acetate	4.29	43	4169	1.804	ug/l #	87
20) Methylene Chloride	4.50	84	14242	5.141	ug/l	99
43) Isopropyl Acetate	8.13	43	45326	8.107	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062221.D
Acq On : 30 Jun 2020 5:48
Operator : JC/MD
Sample : L2806-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-06-062620

Quant Time: Jun 30 06:50:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

