

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062222.D
 Acq On : 30 Jun 2020 6:12
 Operator : JC/MD
 Sample : L2819-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-03-062620

Quant Time: Jun 30 06:52:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	188693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	375800	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	374915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	127357	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	148629	56.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.26%	
35) Dibromofluoromethane	7.55	113	122273	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	10.06	98	491044	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.38	95	154315	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
Target Compounds						
16) Acetone	3.78	43	20284	20.688	ug/l	91
18) Methyl Acetate	4.29	43	4469	2.123	ug/l #	78
20) Methylene Chloride	4.51	84	14928	5.950	ug/l #	95
43) Isopropyl Acetate	8.13	43	38775	7.605	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062222.D
Acq On : 30 Jun 2020 6:12
Operator : JC/MD
Sample : L2819-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-03-062620

Quant Time: Jun 30 06:52:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

