

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062223.D
 Acq On : 30 Jun 2020 6:36
 Operator : JC/MD
 Sample : L3136-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-570-572

Quant Time: Jun 30 08:05:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	195126	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	385274	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	387797	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	133508	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	151224	55.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.46%	
35) Dibromofluoromethane	7.55	113	126342	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
50) Toluene-d8	10.06	98	509395	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
62) 4-Bromofluorobenzene	12.38	95	160599	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
Target Compounds						
16) Acetone	3.78	43	12108	10.201	ug/l	98
18) Methyl Acetate	4.29	43	4709	2.163	ug/l #	70
20) Methylene Chloride	4.50	84	15197	5.854	ug/l	99
43) Isopropyl Acetate	8.13	43	43084	8.242	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062223.D
Acq On : 30 Jun 2020 6:36
Operator : JC/MD
Sample : L3136-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-570-572

Quant Time: Jun 30 08:05:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

