

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062920\
 Data File : VN062224.D
 Acq On : 30 Jun 2020 7:00
 Operator : JC/MD
 Sample : L3136-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-572-574

Manual Integrations
 APPROVED

MMDadoda
 6/30/2020 10:42:20 AM

Quant Time: Jun 30 08:06:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	200812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	404073	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	411014	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	144670	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	160233	56.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.72%	
35) Dibromofluoromethane	7.55	113	128488	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
50) Toluene-d8	10.06	98	538324	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	12.38	95	170357	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
Target Compounds						
16) Acetone	3.78	43	11534	9.136	ug/l	93
18) Methyl Acetate	4.29	43	4205m	1.877	ug/l	
20) Methylene Chloride	4.50	84	17462	6.562	ug/l	92
43) Isopropyl Acetate	8.13	43	44226	8.067	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062920\
Data File : VN062224.D
Acq On : 30 Jun 2020 7:00
Operator : JC/MD
Sample : L3136-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-572-574

Manual Integrations
APPROVED

MMDadoda
6/30/2020 10:42:20 AM

Quant Time: Jun 30 08:06:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

