


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062920\
 Data File : VN062180.D
 Acq On : 29 Jun 2020 9:49
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Title : SW846 8260
 Last Update : Wed Jun 24 15:33:00 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	11087	PASS
75	95	30	60	51.3	27269	PASS
95	95	100	100	100.0	53138	PASS
96	95	5	9	7.4	3945	PASS
173	174	0.00	2	0.9	371	PASS
174	95	50	100	78.1	41520	PASS
175	174	5	9	7.1	2945	PASS
176	174	95	101	97.5	40469	PASS
177	176	5	9	6.7	2702	PASS