

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
 Data File : VN067545.D
 Acq On : 29 Jun 2021 11:37
 Operator : JC/MD
 Sample : VN0629MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0629MBL01

Quant Time: Jun 29 15:21:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	188143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	343428	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	328153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	119174	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	138908	53.947	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.900%
35) Dibromofluoromethane	8.024	113	101910	47.862	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.720%
50) Toluene-d8	10.443	98	467212	54.016	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.040%
62) 4-Bromofluorobenzene	12.734	95	150391	49.777	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
Data File : VN067545.D
Acq On : 29 Jun 2021 11:37
Operator : JC/MD
Sample : VN0629MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0629MBL01

Quant Time: Jun 29 15:21:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

