

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
 Data File : VN067557.D
 Acq On : 29 Jun 2021 16:52
 Operator : JC/MD
 Sample : PB137337TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137337TB

Manual Integrations
 APPROVED

MMDadoda
 7/5/2021 1:51:55 PM

Quant Time: Jun 29 18:07:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	180246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	337060	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	402868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	144507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	136755	55.437	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.880%			
35) Dibromofluoromethane	8.024	113	97753	46.777	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 93.560%			
50) Toluene-d8	10.446	98	490855	57.822	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 115.640%			
62) 4-Bromofluorobenzene	12.733	95	177943	60.009	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 120.020%			
Target Compounds					Qvalue	
16) Acetone	4.301	43	8479	10.421	ug/l	97
20) Methylene Chloride	5.090	84	12340m	5.598	ug/l	
43) Isopropyl Acetate	8.566	43	23916	5.010	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
Data File : VN067557.D
Acq On : 29 Jun 2021 16:52
Operator : JC/MD
Sample : PB137337TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137337TB

Manual Integrations
APPROVED

MMDadoda
7/5/2021 1:51:55 PM

Quant Time: Jun 29 18:07:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

