

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
 Data File : VN067558.D
 Acq On : 29 Jun 2021 17:16
 Operator : JC/MD
 Sample : PB137363TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137363TB

Quant Time: Jun 29 18:07:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	179221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	341800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	408743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	144535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	133585	54.462	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.920%			
35) Dibromofluoromethane	8.021	113	99056	46.743	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 93.480%			
50) Toluene-d8	10.443	98	490489	56.977	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 113.960%			
62) 4-Bromofluorobenzene	12.734	95	179236	59.607	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 119.220%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	9170	11.334	ug/l	98
20) Methylene Chloride	5.090	84	14564	6.683	ug/l	90
25) 2-Butanone	7.354	43	3008	2.411	ug/l	92
43) Isopropyl Acetate	8.566	43	23988	4.956	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
Data File : VN067558.D
Acq On : 29 Jun 2021 17:16
Operator : JC/MD
Sample : PB137363TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137363TB

Quant Time: Jun 29 18:07:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

