

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
 Data File : VN067560.D
 Acq On : 29 Jun 2021 18:05
 Operator : JC/MD
 Sample : M2860-01RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ASHRE

Quant Time: Jun 29 18:23:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	173637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	329635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	392592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	142194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	131518	55.344	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.680%
35) Dibromofluoromethane	8.021	113	95614	46.784	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.560%
50) Toluene-d8	10.443	98	475543	57.280	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.560%
62) 4-Bromofluorobenzene	12.734	95	170837	58.911	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.820%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	9954	12.699	ug/l	93
20) Methylene Chloride	5.095	84	12259	5.779	ug/l	85
25) 2-Butanone	7.354	43	4416	3.653	ug/l	96
43) Isopropyl Acetate	8.566	43	23058	4.940	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
Data File : VN067560.D
Acq On : 29 Jun 2021 18:05
Operator : JC/MD
Sample : M2860-01RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ASHRE

Quant Time: Jun 29 18:23:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

