

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
 Data File : VN067561.D
 Acq On : 29 Jun 2021 18:30
 Operator : JC/MD
 Sample : PB137365TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137365TB

Quant Time: Jun 30 03:40:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	175789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	337827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	407461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	147002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	134199	55.781	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.560%
35) Dibromofluoromethane	8.024	113	97604	46.600	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.200%
50) Toluene-d8	10.446	98	487744	57.325	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.640%
62) 4-Bromofluorobenzene	12.733	95	177219	59.630	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	119.260%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	9015	11.360	ug/l	99
18) Methyl Acetate	4.870	43	5163	2.439	ug/l #	73
20) Methylene Chloride	5.093	84	14916	6.987	ug/l #	81
43) Isopropyl Acetate	8.566	43	23967	5.010	ug/l #	90
59) 2-Hexanone	11.095	43	396395	194.531	ug/l #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
Data File : VN067561.D
Acq On : 29 Jun 2021 18:30
Operator : JC/MD
Sample : PB137365TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137365TB

Quant Time: Jun 30 03:40:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

